

»Retraction paste 2«

de

Produktbeschreibung

Retraction paste 2 ist ein Retraktionsmittel zur Sulkuserweiterung. *Retraction paste 2* wirkt adstringierend auf die Gingiva.

Indikation

Zahnärztliche Fälle, die definitive Abformung präparierter Zähne erfordern z.B. Zahnverlust.

Kontraindikation

Das Material nicht bei Kontaktallergien oder bei bekannten Allergien gegen einen der Inhaltsstoffe verwenden.

Zweckbestimmung

Zur vorübergehenden marginalen Gingivaretraktion und Trockenlegung des Sulkus bei gesundem Parodont

Verwendungsbeschränkung

Nicht anwenden bei exponiertem Knochen, fortgeschrittenen Parodontopathien oder freiliegenden Furkationen.

Vorgesehener Anwender

Zahnarzt

Patientenzielgruppe

Personen, die im Rahmen einer zahnärztlichen Maßnahme behandelt werden.

Empfohlene Anwendung

1. Zahn wie erforderlich präparieren.
2. Die Compule in einen geeigneten Dispenser einsetzen.
3. Spitze der Compule bei Bedarf in erforderliche Position biegen.
4. Eine geringe Menge der Paste auf einen Mischblock ausbringen und verwerfen.
5. Optional: Falls es die klinische Situation erfordert, einen Retraktionsfaden in den Sulkus einlegen.
6. Den Sulkus mit Wasser spülen und trocknen.
7. Paste langsam um die Präparation herum in den Sulkus applizieren, bis der Sulkus vollständig gefüllt ist. Bei Bedarf Kompressionskappen, -hüte oder andere Kompressionsprodukte benutzen.
8. Paste ca. 2 Minuten einwirken lassen, um die Gingivaretraktion zu ermöglichen. Während der Einwirkzeit den behandelten Bereich unbedingt trocken halten.

Vorsicht: Lange Verweildauer der Paste im Mund kann zu Irritationen der Gingiva führen. Mundverweildauer nicht überschreiten.

9. Paste mit einem Wasser/Luft-Gemisch und Absauger vollständig aus dem Sulkus und Mundbereich entfernen.

Hinweis: Paste vollständig entfernen.

10. Reste der Paste können zu Ungenauigkeiten oder zu Verzögerungen der Abbindung des Abformmaterials führen. Gingiva mit Luft trocknen und das Retractionsergebnis kontrollieren. Falls eine stärkere Retraction ge-

wünscht ist, den Vorgang ab Schritt 7 wiederholen. Ggf. zusätzlich einen dünnen Retraktionsfaden verwenden.

11. Abformen

Zeitablauf

Einwirkzeit	2 min
-------------	-------

Hinweise zur Anwendung

Die Compule nach Öffnen des Aluminiumbeutel sofort verwenden. Reste in der Compule verwerfen.

Wechselwirkungen

Reste der Paste können zu Verzögerungen der Abbindung des Abformmaterials führen. Paste mit einem Wasser/Luft-Gemisch und Absauger vollständig aus dem Sulkus und Mundbereich entfernen.

Restrisiken / Nebenwirkungen

Bisher sind keine Nebenwirkungen bekannt. Nicht auszuschließen ist das Restrisiko einer Hypersensibilität gegen Komponenten des Materials.

Warnungen / Vorsichtshinweise

- Für Kinder unzugänglich aufbewahren!
- Kontakt mit der Haut vermeiden! Im Falle eines unbeabsichtigten Hautkontakts die betroffene Stelle sofort gründlich mit Wasser und Seife waschen.
- Kontakt mit den Augen vermeiden! Im Falle eines unbeabsichtigten Augenkontakts sofort gründlich mit viel Wasser spülen und gegebenenfalls einen Arzt konsultieren.
- Herstellerangaben von anderen Produkten, die mit dem Material verwendet werden, beachten.
- Schwerwiegende Vorkommnisse mit diesem Produkt sind dem Hersteller sowie den zuständigen Meldebehörden anzuzeigen.

Leistungsmerkmale des Produkts

Dient der Erweiterung des Sulkus und wirkt adstringierend auf die Gingiva.

Lagerung / Entsorgung

- Bei Raumtemperatur (15 bis 25 °C/59 bis 77 °F) trocken lagern! Nach Ablauf des Haltbarkeitsdatums nicht mehr verwenden.
- Die Compule nach Öffnen des Aluminiumbeutel sofort verwenden. Reste in der Compule verwerfen.
- Unter Beachtung der nationalen Vorschriften entsorgen.

Zusammensetzung

Kaolin, Wasser, Aluminiumchlorid Hexahydrat, Additive.

Handelsformen

REF 9119MDR	5 Compulen à 0,37 g
-------------	---------------------



»Retraction paste 2«

en-us

Product description

Retraction paste 2 is a retraction agent for sulcus widening. *Retraction paste 2* has an astringent effect on the gingiva.

Indications

Dental treatment cases that require final impressions of prepared teeth e.g. after tooth loss.

Contraindications

Do not use the material where there are contact allergies or known allergies to any of the components.

Intended purpose

For temporary marginal gingival retraction and provision of a dry sulcus environment in the case of healthy periodontium.

Limitations of use

Do not use in the case of exposed bone, advanced periodontopathy or exposed furcation defects.

Intended users

Dentist

Patient target group

Persons receiving treatment as part of a dental procedure.

Recommended use

1. Prepare the tooth as required.
2. Place the compule in a suitable dispenser.
3. Bend the tip of the compule into the required position if necessary.
4. Dispense a small amount of paste onto a mixing block and discard.
5. Optional: If needed for the clinical situation, place a retraction cord in the sulcus.
6. Rinse the sulcus with water and then dry the tissue.
7. Place the paste slowly around the preparation in the sulcus until the sulcus is completely filled. Where necessary, use compression caps or other compression products.
8. Leave the paste in place for approx. 02:00 minutes, to enable the gingival retraction. While it is setting, the area being treated must be kept dry.

Caution: Leaving the paste in the mouth for too long can lead to irritations of the gingiva. Do not exceed the intraoral exposure time.

9. Remove the paste completely from the sulcus and mouth area with a mixture of air and water and by using suction.

Note: Completely remove the paste.

10. Residues of the paste can lead to inaccuracies or delays in the setting of the impression material. Dry the gingiva with air and check the retraction result. If a stronger retraction is required, repeat the process from step 7. If necessary, a thin retraction cord can be used in addition.

11. Taking the impression

Timing

Reaction time	02:00 min
---------------	-----------

Notes for use

Use the compule immediately after opening the aluminum bag. Discard any residual material in the compule.

Interactions

Residues of the paste can lead to delays in the setting of the impression material. Remove the paste completely from the sulcus and mouth area with a mixture of air and water and by using suction.

Residual risks / side effects

There are no known side effects to date. The residual risk of hypersensitivity to components in the material cannot be ruled out.

Warnings / precautions

- Keep out of the reach of children!
- Avoid contact with the skin! In the event of accidental skin contact, immediately wash the affected area thoroughly with soap and water.
- Avoid eye contact! In the event of accidental contact with the eyes, immediately rinse thoroughly with plenty of water and consult a physician if necessary.
- Follow the manufacturer's instructions for other products that are used with the material.
- Serious incidents involving this product must be reported to the manufacturer and to the responsible registration authorities.

Performance characteristics of the device

Serves to widen the sulcus and has an astringent effect on the gingiva.

Storage / disposal

- Store in a dry place at room temperature (15 - 25 °C/59 - 77 °F)! Do not use after the expiration date.
- Use the compule immediately after opening the aluminum bag. Discard any residual material in the compule.
- Disposal must comply with national regulations.

Composition

Kaolin clay, water, aluminum chloride hexahydrate, additives.

Packaging

REF 9119MDR	5 compules @ 0.37 g
-------------	---------------------

